

Subject: Reply of CUH on the queries raised by the bidders during the Pre-bid Meeting held on 17/11/2025.

Name of Equipment: Ferroelectric Hysteresis (P-E) Loop Tracer.

Sr. No.	Name of firm	CUH Specification as per NIT	Bidder's queries	Reply of CUH
1.	MICROMA TIC SYSTEMS	(i) Payment Terms	(a) 90% payment against delivery of the instruments and balance 10% after installation, commissioning and submission of PBG. (b) 100% payment after installation, commissioning and submission of PBG within 2 weeks.	No Change
		(ii) Warranty Period	Warranty period for our supplies is 1 year from date of installation. However, if you want extra warranty you have to pay Rs. 8,50,000.00 + GST per year basis.	No Change
		(iii) Regarding UPS	Please note that our system is very sensitive i.e 10 fC and integration with high rating UPS creates noise problems to the users. For this reason, our systems are integrated with the 2-3KVa UPS only. We will provide the 3KVA UPS free of cost.	Recommended
		(vi) Regarding Granite Top Table	Please note that the systems work on High Voltage i.e up to 10,000V as a safety concern we can only provide the wooden table instead of a granite top table.	Recommended