

Subject: Reply of CUH on the queries raised by the bidders during the Pre-bid Meeting held on 17/11/2025.

Name of Equipment: Time Correlated Single Photon Counting (TCSPC) with Low Temperature Photoluminescence (PL) Spectroscopy.

Sr. No.	Name of firm	CUH Specification as per NIT	Bidder's queries	Reply of CUH
1.	Specialise Instruments Marketing Company	Payment terms	90% against delivery and 10% after Installation	No Change
		Delivery	Should be 4 months	No Change

Sr. No.	Name of firm	CUH Specification as per NIT	Bidder's queries	Reply of CUH
2.	Laser Spectra Services India Pvt. Ltd.	Regarding Xe Lamp	Please change the 450W Xe Lamp to 300W Xe Lamp	No change
		Regarding Monochromator	Please change double monochromator to single monochromator in the excitation side	No change